
Replacing the Batteries

1. Insert a coin in the groove on the hattery compartment cover Turn the coin counterclockwise to remove the cover.







3. Align the cover and push 4. Close the battery compartment cover by pushing the cover in with a coin in the groove it down and turning it clockwise until it stops. Groove



Water ingress may damage the battery contacts, causing instrument malfunctions. Note When the battery icon indicates the low power level (), replace both batteries with a brand new set of AAA alkaline batteries (1.5V).

Note Close the cover tightly to seal the battery compartment.

Note Zero-set the instrument after the batteries are replaced. Note Check the expiration dates on batteries before purchase Static images may occasionally appear on the LCD. Such retained pixel charges do not indicate a failure, consume the battery power, or affect the instrument's performance in any way.

Brix scale

Brix represents the weight of sucrose in 100 grams of sucrose solution as percentage by weight. When other dissolved solids are present in the solution. Brix conversion may be applied.

Brix is a measure of the total dissolved solids in a solution and indicates the combined concentration of all soluble substances, such as sugar, salt, protein, and acids,

Automatic Temperature Compensation

The readings are corrected based on the temperature of the prism, within the automatic temperature compensation range.



Note Measurements may fluctuate with hot or cold samples. Wait for approximately 20 seconds to press the START button.

Measurements will stabilize once the instrument acclimates to the sample temperature.

Storage and Maintenance



Store the instrument in a dry place away from direct sunlight. Exposure to humidity and heat may damage the instrument.



Clean and dry the sample stage thoroughly. following the "Cleaning" instructions. Store the unit away from direct sunlight at a stable temperature with as little fluctuation as possible

Repair and Warranty

The instrument is warranted for one year from the date of nurchase

This warranty is void if the instrument shows evidence of the following. Send the included batteries as well if they are still in use.

- · Having been disassembled by unauthorized personnel
- · Water damage or having been dropped
- · Leakage from batteries other than those included with the unit

Repair services are available for a fee after the warranty expires.

Contact an ATAGO authorized service center for service and support

· Damages to the prism and/or sample stage

- · Having been misused and/or operated outside the environmental specifications

Please have the serial number information ready when contacting a service center

Measurement range 0.0 to 53.0% Brix Output Bluetooth Ver 4.0 Maximum communication range: 3 to 7 m Baudrate: 38.4 k/Data length: 8/Prity bit: none RTS/CTS flow control/Stop bit: 1 Transmit data: time, data, Temp 2014/05/30 15: 02: 21,36.6,23.4, 0.1% Brix/0.1 °C ±0.2% Brix/±1 °C Accuracy

CTONO	
Automatic temperature	10 to 100 °C
compensation range	
Ambient temperature range	10 to 40 °C
Sample volume	0.3 ml or more
Measuring time	3 seconds
Power supply	Size AAA alkaline batteries × 2
Battery life	Approx. 10,000 times measurement
	(when an alkaline battery is used)
International Protection class	IP65 Water resistant
Dimensions and weight	55 (W) x 31 (D) x 109 (H)mm,
	100g (main unit only)

The Bluetooth® word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. iPhone, iTunes are trademarks of Apple Inc., registered in the U.S. and other countries. App Store is service mark of Apple Inc.

> The product is in conformity with the requirements of the EMC Directive 2004/108/EC and R&TTE Directive 1999/5/EC. Patent Granted in Japan, United States, Germany, China and Taiwan.

ATAGO CO.,LTD.

Headquarters: The Front Tower Shiba Koen. 23rd Floor 2-6-3 Shiba-koen, Minato-ku. Tokyo 105-0011, Japan TEL: 81-3-3431-1943 overseas@atago.net http://www.atago.net/

ATAGO U.S.A., Inc. TEL: 1-425-637-2107 customerservice@atago-usa.com ATAGO INDIA Instruments Pvt, Ltd. TEL: 91-22-2854-4915 / 4071-3232

customerservice@atago-india.com ATAGO THAILAND Co.,Ltd. TEL: 66-21948727-9 ,66-21171549 customerservice@atago-thailand.com

ATAGO BRASIL Ltda. TEL: 55 16 3913-8400

customerservice@atago-brasil.com

ATAGO ITALIA s.r.l. TEL: 39 02 36557267 customerservice@atago-italia.com

ATAGO CHINA Guangzhou Co., Ltd. TEL: 86-20-38108256

info@atago-china.com

ATAGO RUSSIA Ltd. TEL: 7-812-777-96-96

info@atago-russia.com ATAGO NIGERIA Scientific Co., Ltd. TEL: 234-707-558-1552

atagonigeria@atago.net ATAGO KAZAKHSTAN 📖

TEL: 7-727-257-08-95 info@atago-kazakhstan.com

2108K Printed in Japan

3810i-E04

"Pocket" Refractometer

PAL-1 BLT/ i Cat. No. 3810i

Parts



Instruction Manual

Measurement results, prism temperature and remaining battery charge are displayed

The displayed value is an example.

START button

Press to take measurements and hold down to turn off the display.

START button + ZERO button Press to set the date and time. Be

sure to set the date and time hefore intial use. Battery compartment

Place and remove batteries from here.



Press to perform zero-setting. Press and hold to export data via Bluetooth.

Lanyard hole

····· Contents ····· ◆Main unit ◆Instruction Manual (this book) ◆Calibration Report. ◆AAA batteries ... AAA alkaline batteries are included. Remove the tape from the battery compartment before in-

ATAGO instruments are rigorously inspected to ensure each unit meets the highest standards of quality assurance.

serting the batteries.

Introduction

Thank you for purchasing the instrument, Carefully read and follow all instructions. Keep this manual for future reference.

Safety Instructions

Read and follow all safety instructions before operating the instrument. Failure to comply with the following instructions may result in personal injury or property damage.

⚠ WARNING

- ♦ Ensure safety when handling hazardous materials. Observe precautionary measures and use protective equipment. Be aware of the hazards of such chemicals and emergency response guidelines.
- ♦ ATAGO may not be held liable for any injury or damage arising in connection with handling of hazardous materials during the use of the
- ♦Do not drop the instrument or subject it to strong physical shock.
- ODo not attempt to repair, modify, or disassemble the instrument.

⚠ CAUTION

- Ocarefully read this manual to have basic knowledge of the function of each component.
- ♦ ATAGO is not liable for any loss and damage caused by the measurement and use of this instrument.

- ♦ Some acids may corrode the glass prism and/or metal sample stage. which may cause erroneous measurements.
- ♦Do not use metal tools, such as a spoon, as they may scratch the prism, resulting in erroneous measurements
- ♦ Do not use water above 50°C to rinse the instrument.
- ♦ Only use the specified battery type. Observe proper polarities, properly aligning the anodes and cathodes.
- ♦ Store the instrument away from direct sunlight/heat sources and excessive amounts of dust/debris
- ODo not expose the instrument to a rapid change in ambient temperature.
- ♦ Do not subject the instrument to strong vibration.
- ODo not subject the instrument to extreme cold temperature.
- ♦Do not place the instrument under anything heavy.
- ♦ Loosen the battery compartment cover for air transportation.

⟨International Protection Classification IP65⟩

♦ The instrument is water-resistant, not waterproof, and should not be submerged.

⟨Chemical Resistance of Body Case⟩

♦ The body case is made of resin. Do not expose it to water vapor or solvents. Certain solvents may cause damage to the body case.

FCC Compliance Statement

MODEL: PAL-1 BLT (CONTAINS FCC ID: P00-WC69)

IISA-Federal Communication Commission (ECC)

This device complies with part 15 of the FCC rules. Operation is subject to the following two conditions:

(1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

INFORMATION TO USER:

This equipment has been tested and found to comply with the limit for a Class B digital device, pursuant to Pat 15 of FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy. If not installed and used in accordance with the instructions, it may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on the user is encouraged to try and correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna

- Increase the distance between the equipment and the receiver.
- Connect the equipment to outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

FCC Caution: Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate this equipment.

Canada-Industry Canada

Operation is subject to the following two conditions: (1) this device may not cause interference, and (2) this device must accept any interference. including interference that may cause undesired operation on this device To reduce potential radio interference to other users, the antenna type and its gain should be so chosen that the equivalent isotropically radiated power (EIRP) is not more than that required for successful communication.

Caution: Exposure to Radio Frequency Radiation.

To comply with RSS 102 RF exposure compliance requirements, for mobile configurations, a separation distance of at least 20 cm must be maintained between the antenna of this device and all persons. This device must not be co-located or operating in conjunction with any other antenna or

Zero-setting and Measurement

Date and time setting is required when using for the first time.



Zero-setting Zero-set the instrument at the beginning of each day before use as well as after replacing the batteries.

1. Wipe the nrism clean Apply tap/

3. Press START



distilled water




of water











Zero-set correctly

(82 — Number of measurements stored (Maximum: 100)

Measurement data Full. Oldest measurements will be

overwritten and replaced by new measurements.



Proceed to Measurement

Measurement

1. Wipe the prism clean





2. Apply sample. 3. Press START









Note Initial measurements may fluctuate with hot or cold samples. Wait for the instrument to acclimate to the sample temperature, approximately 20 seconds, to press START). Alternatively, press START multiple times until measurements become stable.

The displayed temperature is that of the prism and may not necessarily match the temperature of the sample.

2. Dry the sample stage completely with dry tissues



Note Do not splash water above 50°C.

stage thoroughly

1. Wipe off the sample, rinse with water, and

wipe off the water to clean the sample

When hot water is necessary to clean off hardened samples, use water-soaked gauze around the prism area and keep hot water away from the body case.

For oily samples:



turn it off, hold down START for more than 2 seconds.







Note When the O-ring on the cover is dirty or damaged, the water resistance may be compromised. Lubricate the O-ring regularly.



Erorr Messages

The following messages alert the user when an operation has failed.



ZERO was pressed with nothing or something other than water on the prism

or an insufficient amount of sample



LO

on the prism The battery is low. TILL "C


The prism temperature is below the temperature range.



The sample is outside the measurement range



Too much light is entering the prism, and the instrument cannot measure accurately. (Shade the sample stage with your hand and take a measurement again.)



The prism temperature is above the temperature range.



The instrument is faulty. (Replace the hatteries Contact ATAGO if this error persists)

data export error.

Date and time setting



* Set date and time before exporting data.

Reset date and time if battery is taken out of the device for an extended period.



Seconds: Fixed 00

